

CERTIFICATE

No. U8 17 09 46515 025

Holder of Certificate: INHECO

Industrial Heating & Cooling GmbH

Fraunhoferstr. 11 82152 Martinsried

GERMANY

Production Facility(ies):

46515

Certification Mark:



Product: Laboratory Equipment

Model(s): Inheco Measurement Plate (IMP)

Parameters: Rated voltage: 5VDC Rated current:

0.2A Protection class: III

Tested UL 61010-1:2012/R:2016-04

CAN/CSA-C22.2 No. 61010-1:2012/U2:2016-04 according to:

1000

The product was voluntarily tested according to the relevant safety requirements noted above. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited certification body.

028-71363929-100 Test report no.:

Date, 2017-09-07

Page 1 of 1

