




America

CERTIFICATE

No. U8 046515 0049 Rev. 00

Holder of Certificate: **INHECO**
Industrial Heating & Cooling GmbH
 Fraunhoferstr. 11
 82152 Martinsried
 GERMANY

Certification Mark:



Product: **Laboratory Equipment
 (Thermal Cycler)**

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL for USA and a Standards Council of Canada ISO/IEC 17065 accredited Certification body for Canada.

Test report no.: 713204190 Rev. 01

Date, 2022-02-21

(Abdul Sabbagh)



America

CERTIFICATE

No. U8 046515 0049 Rev. 00

Model(s): **Multi Tec Control (MTC) devices:**
Thermoshake
Thermoshake AC
Thermoshake AC 180
Teleshake 95
Teleshake AC
Teleshake 95 AC
CPAC Microplate
CPAC Microplate HT 2-TEC
CPAC Ultraflat
CPAC Ultraflat HT 2-TEC
CPAC Ultraflat HT 2-TEC Down Vent
CPLC Ultraflat MTC
CPLC Ultraflat HT 2-TEC
HeatPAC
Heated Lid

Accessory for use with MTC devices:
Heat Exchanger Liquid cooling small

Brand Name: **inheco**

Tested according to: UL 61010-1:2012/R:2016-04
 UL 61010-2-010:2015
 UL 61010-2-101:2015
 CAN/CSA-C22.2 No. 61010-1:2012/U2:2016-04
 CAN/CSA-C22.2 No. 61010-2-010:2015
 CAN/CSA-C22.2 No. 61010-2-101:2015

Parameters:
Thermoshake; Thermoshake AC; Thermoshake AC 180; Teleshake 95 AC:

Rated voltage: 24 Vdc
 Rated current: 5.3 A
 Protection class: III

Teleshake 95; Heated Lid:

Rated voltage: 24 Vdc
 Rated current: 4.8 A
 Protection class: III

Teleshake AC:

Rated voltage: 24 Vdc
 Rated current: 1.5 A
 Protection class: III



America

CERTIFICATE

No. U8 046515 0049 Rev. 00

CPAC Microplate; CPLC Ultraflat MTC:

Rated voltage: 12 Vdc
 Rated current: 5.0 A
 Protection class: III

CPAC Microplate HT 2-TEC:

Rated voltage: 24 Vdc
 Rated current: 4.7 A
 Protection class: III

CPAC Ultraflat:

Rated voltage: 12 Vdc
 Rated current: 4.5 A
 Protection class: III

CPAC Ultraflat HT 2-TEC; CPAC Ultraflat HT 2-TEC Down Vent:

Rated voltage: 24 Vdc
 Rated current: 4.3 A
 Protection class: III

CPLC Ultraflat HT 2-TEC:

Rated voltage: 24 Vdc
 Rated current: 4.08 A
 Protection class: III

HeatPAC:

Rated voltage: 24 Vdc
 Rated current: 4.16 A
 Protection class: III

Heat Exchanger Liquid cooling small:

Rated voltage: 12 Vdc
 Rated current: 50 W
 Protection class: III