

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

## CB TEST CERTIFICATE

Product	Laboratory Equipment (Cell Incubator for Lab Automation)
Name and address of the applicant	<b>INHECO</b> <b>Industrial Heating &amp; Cooling GmbH</b> Fraunhoferstr. 11 82152 Martinsried GERMANY
Name and address of the manufacturer	INHECO Industrial Heating & Cooling GmbH Fraunhoferstr. 11, 82152 Martinsried, GERMANY
Name and address of the factory	INHECO Industrial Heating & Cooling GmbH Fraunhoferstr. 11, 82152 Martinsried, GERMANY
Ratings and principal characteristics	Rated voltage: 24 Vdc  Rated power: 130 W  Protection class: III
Trade mark (if any)	inheco
Model/type Ref.	SCILA CO2 MP-4
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010 IEC 61010-1:2010/AMD1:2016 IEC 61010-2-010:2019 IEC 61010-2-081:2019
as shown in the Test Report Ref. No. which forms part of this certificate	713220620

Page 1 of 1

This CB Test Certificate is issued by the National Certification Body

CB 046515 0045 Rev. 00

Date, 2021-10-18



( Abdul Sabbagh )

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service