



Ref. Certif. No.

BE-33710/M1

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

SCILA Incubator for Lab Automation

Name and address of the applicant

INHECO GmbH
Fraunhoferstrasse 11, 82152, Martinsried, Germany

Name and address of the manufacturer

INHECO GmbH
Fraunhoferstrasse 11, 82152, Martinsried, Germany

Name and address of the factory

INHECO GmbH
Fraunhoferstrasse 11, 82152, Martinsried, Germany


Note: When more than one factory, please report on page 2

☐ Additional Information on page 2

Ratings and principal characteristics

24 VDC; 130 W

Trademark (if any)



Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

SCILA CO2 MP-4, SCILA CO2 DWP-2

Additional information (if necessary may also be reported on page 2)

Re-Issue of BE-33710 dated 2018-12-14 due to technical change : change in list of critical components. For further details please refer to test report.

☐ Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010
IEC 61010-2-010:2014
IEC 61010-2-081:2015
National Differences:
CA, US

N0E50002, N0E50003, N0E50004, N0E50005

As shown in the Test Report Ref. No. which forms part of this Certificate

This CB Test Certificate is issued by the National Certification Body

SGS Belgium NV - Division SGS CEBEC
Riverside Business Park
Avenue Internationalelaan 55, Building D
B-1070 Brussels, Belgium

SGS



Date: 2020-01-21

Signature: Calogero Lana