



Product Service

# Attestation of Conformity

No. N8A 046515 0048 Rev. 00

**Holder of Certificate:** **INHECO**  
**Industrial Heating & Cooling GmbH**  
Fraunhoferstr. 11  
82152 Martinsried  
GERMANY

**Product:** **Laboratory Equipment**  
**(Cell Incubator for Lab Automation)**

**Brand:** **inheco**

**Model(s):** **SCILA CO2 MP-4**

**Parameters:** Rated voltage: 24 Vdc  
Rated power: 130 W  
Protection class: III

**Tested according to:** EN 61010-1:2010/A1:2019  
EN IEC 61010-2-010:2020  
EN IEC 61010-2-081:2020

This Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 2014/35/EU relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification. For details see: [www.tuvsud.com/ps-cert](http://www.tuvsud.com/ps-cert)

**Test report no.:** 713220620

**Date,** 2021-10-18



( Abdul Sabbagh )