



Product Service

CERTIFICATE

No. Z1 046515 0047 Rev. 00

Holder of Certificate: **INHECO**
Industrial Heating & Cooling GmbH
Fraunhoferstr. 11
82152 Martinsried
GERMANY

Certification Mark:



Product: **Laboratory Equipment**
(Cell Incubator for Lab Automation)

Model(s): **SCILA CO2 MP-4**

Parameters: Rated voltage: 24 Vdc
Rated power: 130 W
Protection class: III

Brand Name: inheco

Tested according to: EN 61010-1:2010/A1:2019
EN IEC 61010-2-010:2020
EN IEC 61010-2-081:2020

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition, the certification holder must not transfer the certificate to third parties. This certificate is valid until the listed date, unless it is cancelled earlier. All applicable requirements of the testing and certification regulations of TÜV SÜD Group have to be complied. For details see: www.tuvsud.com/ps-cert

Test report no.: 713220620
Valid until: 2026-10-14

Date, 2021-10-19

(Abdul Sabbagh)