




America

CERTIFICATE

No. U8 046515 0046 Rev. 00

Holder of Certificate: **INHECO**
Industrial Heating & Cooling GmbH
Fraunhoferstr. 11
82152 Martinsried
GERMANY

Certification Mark:



Product: **Laboratory Equipment**
(Cell Incubator for Lab Automation)

Brand Name: **inheco**

Model(s): **SCILA CO2 MP-4**

Parameters: Rated voltage: 24 Vdc
Rated power: 130 W
Protection class: III

Tested according to: UL 61010-1:2012/R:2019-07
CAN/CSA-C22.2 No. 61010-1:2012/A1:2018-11
UL 61010-2-010:2019
CAN/CSA-C22.2 No. 61010-2-010:2019
UL 61010-2-081:2019
CSA C22.2 No. 61010-2-081:2019

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

Test report no.: 713220620

Date, 2021-10-18

(Abdul Sabbagh)