




America

CERTIFICATE

No. U8 046515 0040 Rev. 00

Holder of Certificate: **INHECO**
Industrial Heating & Cooling GmbH
Fraunhoferstr. 11
82152 Martinsried
GERMANY

Certification Mark:



Product: **Laboratory Equipment**
(Thermal Cycler)

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

Test report no.: 028-713186800-000

Date, 2020-08-17

(Abdul Sabbagh)



America

CERTIFICATE

No. U8 046515 0040 Rev. 00

Model(s): **Multi Tec Control (MTC) devices:**

Thermoshake
Thermoshake AC
Thermoshake AC 180
Teleshake 95
Teleshake AC
Teleshake 95 AC
CPAC Microplate
CPAC Microplate HT 2-TEC
CPAC Ultraflat
CPAC Ultraflat HT 2-TEC
CPAC Ultraflat HT 2-TEC Down Vent
CPLC Ultraflat MTC
HeatPAC
Heated Lid

Brand Name: **inheco**

Tested according to:

UL 61010-1:2012/R:2016-04
 UL 61010-2-010:2015
 UL 61010-2-101:2015
 CAN/CSA-C22.2 No. 61010-1:2012/U2:2016-04
 CAN/CSA-C22.2 No. 61010-2-010:2015
 CAN/CSA-C22.2 No. 61010-2-101:2015

Production Facility(ies): 046515

Parameters: **Thermoshake; Thermoshake AC; Thermoshake AC 180; Teleshake 95 AC**

Rated voltage: 24 Vdc
 Rated current: 5.3 A
 Protection class: III

Teleshake 95; Heated Lid

Rated voltage: 24 Vdc
 Rated current: 4.8 A
 Protection class: III



CERTIFICATE

No. U8 046515 0040 Rev. 00

Teleshake AC

Rated voltage: 24 Vdc
 Rated current: 1.5 A
 Protection class: III

CPAC Microplate

Rated voltage: 12 Vdc
 Rated current: 5.0 A
 Protection class: III

CPAC Microplate HT 2-TEC

Rated voltage: 24 Vdc
 Rated current: 4.7 A
 Protection class: III

CPAC Ultraflat

Rated voltage: 12 Vdc
 Rated current: 4.5 A
 Protection class: III

CPAC Ultraflat HT 2-TEC; CPAC Ultraflat HT 2-TEC Down Vent

Rated voltage: 24 Vdc
 Rated current: 4.3 A
 Protection class: III

CPLC Ultraflat MTC

Rated voltage: 12 Vdc
 Rated current: 5.0 A
 Protection class: III

HeatPAC

Rated voltage: 24 Vdc
 Rated current: 4.16 A
 Protection class: III