



America

CERTIFICATE

No. U8 17 09 46515 028

Holder of Certificate: **INHECO**
Industrial Heating & Cooling GmbH

Fraunhoferstr. 11
82152 Martinsried
GERMANY

Certification Mark:



Product: **Laboratory Equipment**

The product was voluntarily tested according to the relevant safety requirements noted above. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited certification body.

Test report no.: 028-71363947-200

Date, 2017-09-28

Page 1 of 2





America

CERTIFICATE

No. U8 17 09 46515 028

Model(s): Incubator MP x nnn;
 Incubator DWP x nnn;
 Incubator Shaker MP x nnn;
 Incubator Shaker DWP x nnn,
 where
 x: can be blank or U, or V, or PCR and represents
 the bottom shape;
 nnn: can be blank or 24, or 96, or 384 and represents
 the probe format.

Brand Name: inheco

Parameters:

Incubator MP x nnn	
Rated voltage:	24 Vdc
Rated power:	30 W
Protection class:	III
Incubator DWP x nnn	
Rated voltage:	24 Vdc
Rated power:	45 W
Protection class:	III
Incubator Shaker MP x nnn	
Rated voltage:	24 Vdc
Rated power:	50 W
Protection class:	III
Incubator Shaker DWP x nnn	
Rated voltage:	24 Vdc
Rated power:	65 W
Protection class:	III

Tested according to:

UL 61010-1:2012/R:2016-04
 UL 61010-2-010:2015-01
 UL 61010-2-101:2015-08
 CAN/CSA-C22.2 No. 61010-1:2012/U2:2016-04
 CAN/CSA-C22.2 No. 61010-2-010:2015-01
 CAN/CSA-C22.2 No. 61010-2-101:2015-10

Production Facility(ies): 46515